



THE AMERICAN ASSOCIATION FOR
LABORATORY ACCREDITATION

ACCREDITED LABORATORY

A2LA has accredited

MOREHOUSE INSTRUMENT CO., INC.

York, PA

for technical competence in the field of

Calibration

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2005 *General Requirements for the Competence of Testing and Calibration Laboratories*. This laboratory also meets the requirements of ANSI/NCSL Z540-1-1994 and any additional program requirements in the field of calibration. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (*refer to joint ISO-ILAC-IAF Communiqué dated 18 June 2005*).



Presented this 9th day of April 2008.

A handwritten signature in cursive script, reading "Peter Abney".

President
For the Accreditation Council
Certificate Number 1398.01
Valid to April 30, 2010

For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Accreditation.

SCOPE OF ACCREDITATION TO ISO/IEC 17025:2005
& ANSI/NCSL Z540-1-1994

MOREHOUSE INSTRUMENT CO., INC.
1742 Sixth Avenue
York, PA 17403-2675
Harry E. Zumbrun Phone: 717 843 0081

CALIBRATION

Valid To: April 30, 2010

Certificate Number: 1398.01

In recognition of the successful completion of the A2LA evaluation process, accreditation is granted to this laboratory to perform the following calibrations¹:

I. Mechanical

Parameter	Range	Best Uncertainty ² (±)	Comments
Force – Tension/Compression	(0 to 120) lbf (0 to 1100) lbf (0 to 12 000) lbf (0 to 120 000) lbf	0.0028% of applied force 0.0033 % of applied force 0.0026 % of applied force 0.0021 % of applied force	Force calibration including ASTM E74 and ISO 376 using primary standards (dead weights)
	(100 000 to 1 000 000) lbf	19 lbf	Secondary standards

¹ This laboratory offers commercial calibration service.

² “Best Uncertainty” is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards of nearly ideal measuring equipment. Best uncertainties represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of $k = 2$. The best uncertainty of a specific calibration performed by the laboratory may be greater than the best uncertainty due to the behavior of the customer’s device and to influences from the circumstances of the specific calibration.